

VERIFICATION OF CALIBRATION REPORT

NOV 16 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N: 130135 ✓

Name: Trp. J. Santora

ID# 2276 ✓

Date: 11/9/10 ✓

A

Agency: Alaska State Troopers

Phone # 656-1233

Instrument Location: Galena Police Department

B

Alco S/N: 78855 ✓

Target Value: .081 ✓

High Pressure: 1050

Alco Test Values

.081 ✓

1st Alco

.080 ✓

2nd Alco

Signature

J. Santora

(OVER)

DMB
12/16/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this 23rd day of DEC, 2010.

(Notary Seal Stamp)

Daniel L. Nicholas

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130135

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135

NOVEMBER 09, 2010

OPERATOR'S NAME:

SANTORA, JEANNINE L

OPERATOR'S NUMBER: 2276

SUBJECT'S LAST NAME:

VERIFICATION

SUBJECT'S FIRST NAME/MI :
OF CAL

O.L. #: 0000000

DEPT/AGENCY: GALO

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .001

ALCO S/N: 78855

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.96 in
ALCO TARGET .001 11:42
BLANK TEST .000 11:43
INTERNAL STANDARD VERIFIED 11:43
ALCO TO 29.96 in .001 11:44
BLANK TEST .000 11:45
SUBJECT SAMPLE .000 11:45
BLANK TEST .000 11:46
ALCO TO 29.96 in .000 11:46
BLANK TEST .000 11:47

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135

NOVEMBER 09, 2010

TIME 11:48

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 29.96 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL MNOPQRSTUVWXYZ \]^_`abcde fghijklmno
pqrstuvwxyz{|}~